

PHOTONIC SENSOR MONITOR

Laboratory Tabletop for Photonic Process Control



Photonic Sensor Monitor

The two channel **Photonic Sensor Monitor** is used in photonic laboratories. It reads sensor signals generated by photodiodes as well as signals from amplified probes (0...2.5V). The unit displays the calibrated intensity values and calculates radiation doses. The included PC software SensorView 1.2 allows data logging and computer based scientific data analysis. Accessories are two sensor cables and a power supply (110...220V). NIST or PTB calibrated sensor inputs are available on request.

The unit is also available as a compact OEM module for industrial use.

Features Overview

Inputs	two channel input, all common photonic sensors and photodiodes, NIST or PTB traceable input calibration on request
Software	SensorView 1.2 for data logging and scientific data evaluation
Outputs	three programmable free floating relay terminals to control minimum and maximum intensity values and dose values; complex logic operations programmable (see page 4)

Specifications	Value	Unit
Numbers of input channels	2	-
Data interface	USB/RS232	-
Numbers of free floating relay	3	-
Dimensions (HxWxD)	234x95X197	mm
Degree of protection	IP40	-
Operating temperature	0...+70	°C
Storage temperature	-25... +85	°C
Power supply	12... 24	V _{DC}
Power consumption (24V)	0,4	W
Weight	1,22	kg

PHOTONIC SENSOR MONITOR

Laboratory Tabletop for Photonic Process Control



The Inputs and Outputs of the Photonic Sensor Monitors

Sensor Inputs



The sensor input terminals can be connected to photodiodes or pre-amplified sensors


Value and status information display



The full-programmable two row display shows e.g.

- Radiation, dose and status of the the relay
- error messages like overrange

Data evaluation



The software SensorView 1.2 allows a comfortable data evaluation. Alternatively a USB/RS232 data export to other scientific evaluation softwares such as Origin or Excel is possible.

Relay Outputs



The three free floating relay allow simple and complex process control. Examples:

- activation if a programmable threshold is exceeded or undershot
- activation if a programmable dose is reached
- logic combination of the two sensor inputs such as controll of multi-step irradiation processes, e.g. swith off lamp #1 if dose #1 is reached and switch on lamp #2 and start of a conveyor belt

PHOTONIC SENSOR MONITOR

Laboratory Tabletop for Photonic Process Control




Basic Functions

Radiation Measurement


Parallel measurements of two lamp outputs or radiation parts of one lamp can be realized (e.g. UVA and UVB). In the first display example the intensity I1 at sensor 1 and the intensity I2 at sensor 2 is displayed. The second picture shows intensity and dose (time integration of the intensity).

If the data port is activated the complete relevant information (intensities, doses, error messages and state of relays and dose measurements) is transferred to a PC.




Measurement and Lamp Control

At exceeding or falling below a configurable intensity threshold or reaching an irradiation dose the lamp can be switched off or changed over to another lamp.




Measurement and Alarm

In the example an alarm is given if the percentaged lamp power falls below a configurable threshold. Further two relays can be used for other functions (switching of pumps, shutters etc.).



Transport Control of irradiated Goods

Measurement of the dose at irradiated goods and activation of the belt transport. The hold times of the relays are variable therefore the transport distance can be adjusted with the hold time. In the example a second relay is giving an alarm if the intensity falls below the threshold. The third relay can be used for information from a second sensor or for a logic combination with one of the other relays (e.g. transport if dose threshold is exceeded and intensity is higher than a minimum value at the same time).



PHOTONIC SENSOR MONITOR

Laboratory Tabletop for Photonic Process Control




Advanced Use in Process Automation


Automated Dose Measurement

The measurement of irradiation doses can be done manually or subjected to automation conditions.

In the first example the dose measurement is started at exceeding a critical intensity. If the intensity falls below the threshold the integration is interrupted and the dose stays constant. While exceeding the threshold again, the integration is continued.




In the second example the dose measurement is finished with falling under the intensity threshold. The reset delay keeps the value on the display. At exceeding of the threshold a new dose is generated. The generation of single doses is used if the dose stop condition is activating a pump or a transport of a good (see below). For each irradiated good or segment a dose is calculated.



Relay Configuration


Three relays can be configured for controlling different functions activated by various configurable process conditions. In the simplest use the relays activate at falling under or exceeding a critical threshold of a selectable measure.

A warmup delay can be implemented to avoid false reports at the start-up process. Additionally it may be reasonable to ignore a short malfunction and only to consider a longer malfunction by using a relay activation delay. Hysteresis parameters can be set for values that are alternating around the threshold.



Dose Measurement Indication

Each relay can be associated with dose functions. Running dose measurements can be indicated by an activated relay. There is no difference if the dose measurement is operated manually or under automated conditions.



Dose Limit Indication

At the dose limit indication the relay is activated if the dose measurement is finished. With the hold time the time of the relay activation is set.

